

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD072518\
 Data File : VD059600.D
 Acq On : 25 Jul 2018 20:26
 Operator : VA/AP
 Sample : J4026-08RE
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DUP-0492-180716RE

Quant Time: Jul 26 05:36:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072418S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 25 07:10:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	25229	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.67	114	36993	50.00	ug/l	0.01
63) Chlorobenzene-d5	10.68	117	34030	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	12.90	152	16604	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.99	65	27264	85.96	ug/l	0.01
Spiked Amount			Recovery	=	171.92%	
35) Dibromofluoromethane	5.52	113	19026	59.96	ug/l	0.01
Spiked Amount			Recovery	=	119.92%	
50) Toluene-d8	8.69	98	38187	53.05	ug/l	0.01
Spiked Amount			Recovery	=	106.10%	
62) 4-Bromofluorobenzene	11.92	95	17936	55.32	ug/l	0.01
Spiked Amount			Recovery	=	110.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.59	43	10518	219.59	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD072518\
Data File : VD059600.D
Acq On : 25 Jul 2018 20:26
Operator : VA/AP
Sample : J4026-08RE
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DUP-0492-180716RE

Quant Time: Jul 26 05:36:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D072418S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 25 07:10:33 2018
Response via : Initial Calibration

