

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072623\
 Data File : VD076810.D
 Acq On : 26 Jul 2023 17:15
 Operator : KP/SY
 Sample : 03769-05
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-P15A

Quant Time: Jul 27 04:24:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	197199	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	403328	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	368097	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	159881	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	106444	54.193	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.380%	
35) Dibromofluoromethane	7.805	113	132531	52.290	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.580%	
50) Toluene-d8	10.269	98	489904	54.031	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 108.060%	
62) 4-Bromofluorobenzene	12.575	95	142123	48.453	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 96.900%	
Target Compounds						
					Qvalue	
17) Carbon Disulfide	4.269	76	4846	1.065	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072623\
Data File : VD076810.D
Acq On : 26 Jul 2023 17:15
Operator : KP/SY
Sample : 03769-05
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-P15A

Quant Time: Jul 27 04:24:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

