


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD072720\
Data File : VD066061.D
Acq On : 27 Jul 2020 09:11
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072120S.M
Title : SW846 8260
Last Update : Tue Jul 21 12:34:13 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.6	24749	PASS
75	95	30	60	51.3	68200	PASS
95	95	100	100	100.0	132992	PASS
96	95	5	9	6.7	8897	PASS
173	174	0.00	2	0.5	598	PASS
174	95	50	100	94.2	125301	PASS
175	174	5	9	8.8	11045	PASS
176	174	95	101	96.4	120805	PASS
177	176	5	9	7.5	9041	PASS