

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072723\
 Data File : VD076833.D
 Acq On : 27 Jul 2023 14:44
 Operator : KP/SY
 Sample : 03769-13
 Misc : 5.10G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-P31A

Quant Time: Jul 27 15:04:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	185459	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	386276	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	355937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	147157	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	102448	55.460	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.920%
35) Dibromofluoromethane	7.804	113	126502	52.114	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.220%
50) Toluene-d8	10.269	98	473839	54.567	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.140%
62) 4-Bromofluorobenzene	12.575	95	134414	47.848	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.700%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072723\
Data File : VD076833.D
Acq On : 27 Jul 2023 14:44
Operator : KP/SY
Sample : 03769-13
Misc : 5.10G/5.0ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-P31A

Quant Time: Jul 27 15:04:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD076833.D
Acq: 27 Jul 2023 14:44




Instrument :
MSVOA_D
ClientSampleId :
B-P31A


Tgt Ion:168 Resp: 185459
Ion Ratio Lower Upper
168 100
99 57.8 36.7 55.1#



#33
1,2-Dichloroethane-d4
Concen: 55.460 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. -0.000 min
Lab File: VD076833.D
Acq: 27 Jul 2023 14:44

Tgt Ion: 65 Resp: 102448
Ion Ratio Lower Upper
65 100
67 56.2 0.0 115.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.775 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD076833.D
Acq: 27 Jul 2023 14:44




Instrument :
MSVOA_D
ClientSampleId :
B-P31A


Tgt Ion:114 Resp: 386276
Ion Ratio Lower Upper
114 100
63 17.3 0.0 33.4
88 13.8 0.0 26.2



#35
Dibromofluoromethane
Concen: 52.114 ug/l
RT: 7.804 min Scan# 1057
Delta R.T. -0.001 min
Lab File: VD076833.D
Acq: 27 Jul 2023 14:44

Tgt Ion:113 Resp: 126502
Ion Ratio Lower Upper
113 100
111 101.0 84.7 127.1
192 17.5 16.8 25.2









#50
Toluene-d8
Concen: 54.567 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD076833.D
Acq: 27 Jul 2023 14:44




Instrument :
MSVOA_D
ClientSampleId :
B-P31A


Tgt Ion: 98 Resp: 473839
Ion Ratio Lower Upper
98 100
100 64.3 50.3 75.5



#62
4-Bromofluorobenzene
Concen: 47.848 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD076833.D
Acq: 27 Jul 2023 14:44

Tgt Ion: 95 Resp: 134414
Ion Ratio Lower Upper
95 100
174 86.1 0.0 201.8
176 82.1 0.0 191.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 117
Delta R.T. -0.000 min
Lab File: VD076833.D
Acq: 27 Jul 2023 14:44

Instrument :
MSVOA_D
ClientSampleId :
B-P31A

Tgt Ion:117 Resp: 355937
Ion Ratio Lower Upper
117 100
82 52.1 38.3 57.5
119 32.2 25.0 37.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.006 min
Lab File: VD076833.D
Acq: 27 Jul 2023 14:44

Tgt Ion:152 Resp: 147157
Ion Ratio Lower Upper
152 100
115 59.0 27.1 81.3
150 154.5 0.0 353.6

