

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072821\
 Data File : VD069886.D
 Acq On : 28 Jul 2021 14:00
 Operator : VA/SY
 Sample : M3204-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-01-072721

Quant Time: Jul 29 03:59:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	51664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	87601	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	93522	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	38369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	48404	81.463	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 162.920%	
35) Dibromofluoromethane	7.908	113	29661	53.711	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.420%	
50) Toluene-d8	10.338	98	112072	52.631	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.260%	
62) 4-Bromofluorobenzene	12.626	95	37229	52.314	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.620%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.355	83	3199	2.955	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072821\
Data File : VD069886.D
Acq On : 28 Jul 2021 14:00
Operator : VA/SY
Sample : M3204-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-01-072721

Quant Time: Jul 29 03:59:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

