

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072821\
 Data File : VD069889.D
 Acq On : 28 Jul 2021 15:25
 Operator : VA/SY
 Sample : M3202-01
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CL-02-072721

Quant Time: Jul 29 03:59:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	15088	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	23779	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	22679	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	7434	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	12307	70.923	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	141.840%
35) Dibromofluoromethane	7.908	113	8095	54.002	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.000%
50) Toluene-d8	10.337	98	29796	51.549	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.100%
62) 4-Bromofluorobenzene	12.620	95	7591	39.296	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072821\
Data File : VD069889.D
Acq On : 28 Jul 2021 15:25
Operator : VA/SY
Sample : M3202-01
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CL-02-072721

Quant Time: Jul 29 03:59:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

