

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072821\
 Data File : VD069897.D
 Acq On : 28 Jul 2021 19:15
 Operator : VA/SY
 Sample : M3216-09
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20210593-AMENDED-COM-7

Quant Time: Jul 29 04:02:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	620022	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1279961	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1167936	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	433730	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	427029	59.885	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.780%
35) Dibromofluoromethane	7.909	113	348222	43.156	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.320%
50) Toluene-d8	10.338	98	1627193	52.299	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.600%
62) 4-Bromofluorobenzene	12.626	95	527769	50.756	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.520%
Target Compounds						
						Qvalue
16) Acetone	4.162	43	29672	23.035	ug/l	93
17) Carbon Disulfide	4.409	76	157262	6.200	ug/l	100
84) 1,2,4-Trimethylbenzene	13.255	105	27038	1.028	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072821\
Data File : VD069897.D
Acq On : 28 Jul 2021 19:15
Operator : VA/SY
Sample : M3216-09
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20210593-AMENDED-COM-7

Quant Time: Jul 29 04:02:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

