

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072821\
 Data File : VD069899.D
 Acq On : 28 Jul 2021 20:12
 Operator : VA/SY
 Sample : M3206-01
 Misc : 8.83G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20210428-14

Quant Time: Jul 29 04:03:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	714957	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1405463	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1268466	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	531173	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	500556	60.875	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 121.760%			
35) Dibromofluoromethane	7.909	113	457896	51.681	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 103.360%			
50) Toluene-d8	10.338	98	1766632	51.711	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 103.420%			
62) 4-Bromofluorobenzene	12.626	95	623529	54.611	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 109.220%			
Target Compounds					Qvalue	
16) Acetone	4.156	43	114285	76.939	ug/l	94
17) Carbon Disulfide	4.409	76	562816	19.243	ug/l	97
25) 2-Butanone	7.214	43	32264	16.017	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072821\
Data File : VD069899.D
Acq On : 28 Jul 2021 20:12
Operator : VA/SY
Sample : M3206-01
Misc : 8.83G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20210428-14

Quant Time: Jul 29 04:03:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

