

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072823\
 Data File : VD076847.D
 Acq On : 28 Jul 2023 11:36
 Operator : KP/SY
 Sample : VD0728SBL01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0728SBL01

Quant Time: Jul 28 23:00:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	202317	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	392273	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	354478	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	151770	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	99037	49.146	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.300%
35) Dibromofluoromethane	7.805	113	123188	49.973	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.940%
50) Toluene-d8	10.269	98	467636	53.029	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.060%
62) 4-Bromofluorobenzene	12.575	95	132098	46.305	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072823\
Data File : VD076847.D
Acq On : 28 Jul 2023 11:36
Operator : KP/SY
Sample : VD0728SBL01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0728SBL01

Quant Time: Jul 28 23:00:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

