

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072823\
 Data File : VD076851.D
 Acq On : 28 Jul 2023 13:29
 Operator : KP/SY
 Sample : 03770-04
 Misc : 5.06G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DEL-2-RP

Quant Time: Jul 28 23:34:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	191770	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	362259	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	333659	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	139980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	93160	48.773	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.540%
35) Dibromofluoromethane	7.804	113	116063	50.984	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.960%
50) Toluene-d8	10.269	98	442421	54.326	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.660%
62) 4-Bromofluorobenzene	12.575	95	123706	46.956	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	93.920%
Target Compounds						
16) Acetone	4.022	43	27891	61.142	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072823\
Data File : VD076851.D
Acq On : 28 Jul 2023 13:29
Operator : KP/SY
Sample : 03770-04
Misc : 5.06G/5.0ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DEL-2-RP

Quant Time: Jul 28 23:34:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

