

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072823\
 Data File : VD076859.D
 Acq On : 28 Jul 2023 17:20
 Operator : KP/SY
 Sample : 03810-06
 Misc : 5.03G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-P7B

Quant Time: Jul 28 23:06:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	172931	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	334955	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	315453	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	132818	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	91010	52.838	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.680%	
35) Dibromofluoromethane	7.805	113	106784	50.732	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.460%	
50) Toluene-d8	10.269	98	407978	54.181	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 108.360%	
62) 4-Bromofluorobenzene	12.575	95	114283	46.915	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 93.840%	
Target Compounds						
						Qvalue
16) Acetone	4.022	43	9477	23.038	ug/l	92
17) Carbon Disulfide	4.252	76	6139	1.538	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072823\
Data File : VD076859.D
Acq On : 28 Jul 2023 17:20
Operator : KP/SY
Sample : 03810-06
Misc : 5.03G/5.0ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-P7B

Quant Time: Jul 28 23:06:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

