

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072823\
 Data File : VD076860.D
 Acq On : 28 Jul 2023 17:49
 Operator : KP/SY
 Sample : 03810-07
 Misc : 5.05G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-P15B

Quant Time: Jul 29 00:17:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	122460	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	243100	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	231978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	101244	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	71458	58.585	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.160%
35) Dibromofluoromethane	7.798	113	79872	52.284	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.560%
50) Toluene-d8	10.269	98	290248	53.110	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.220%
62) 4-Bromofluorobenzene	12.575	95	88980	50.330	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	100.660%
Target Compounds						
16) Acetone	4.016	43	10479	35.973	ug/l	# 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072823\
Data File : VD076860.D
Acq On : 28 Jul 2023 17:49
Operator : KP/SY
Sample : 03810-07
Misc : 5.05G/5.0ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-P15B

Quant Time: Jul 29 00:17:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

