

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072823\
 Data File : VD076857.D
 Acq On : 28 Jul 2023 16:22
 Operator : KP/SY
 Sample : 03810-04
 Misc : 5.08G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-P16B

Quant Time: Jul 28 23:05:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	179973	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	338456	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	320803	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	140402	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	98321	54.849	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.700%
35) Dibromofluoromethane	7.804	113	108561	51.042	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.080%
50) Toluene-d8	10.269	98	418664	55.025	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	110.040%
62) 4-Bromofluorobenzene	12.575	95	119765	48.657	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.320%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072823\
Data File : VD076857.D
Acq On : 28 Jul 2023 16:22
Operator : KP/SY
Sample : 03810-04
Misc : 5.08G/5.0ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-P16B

Quant Time: Jul 28 23:05:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD076857.D
Acq: 28 Jul 2023 16:22




Instrument :
MSVOA_D
ClientSampleId :
B-P16B


Tgt Ion:168 Resp: 179973
Ion Ratio Lower Upper
168 100
99 49.4 36.7 55.1



#33
1,2-Dichloroethane-d4
Concen: 54.849 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. -0.000 min
Lab File: VD076857.D
Acq: 28 Jul 2023 16:22

Tgt Ion: 65 Resp: 98321
Ion Ratio Lower Upper
65 100
67 52.8 0.0 115.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1

Delta R.T. -0.000 min

Lab File: VD076857.D

Acq: 28 Jul 2023 16:22

Instrument :

MSVOA_D

ClientSampleId :

B-P16B





Tgt Ion:114 Resp: 338456

Ion Ratio Lower Upper

114 100

63 18.4 0.0 33.4

88 14.5 0.0 26.2



#35

Dibromofluoromethane

Concen: 51.042 ug/l

RT: 7.804 min Scan# 1057

Delta R.T. -0.001 min

Lab File: VD076857.D

Acq: 28 Jul 2023 16:22




Tgt Ion:113 Resp: 108561


Ion Ratio Lower Upper

113 100

111 104.1 84.7 127.1

192 18.3 16.8 25.2









#50
Toluene-d8
Concen: 55.025 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD076857.D
Acq: 28 Jul 2023 16:22




Instrument :
MSVOA_D
ClientSampleId :
B-P16B


Tgt Ion: 98 Resp: 418664
Ion Ratio Lower Upper
98 100
100 65.1 50.3 75.5



#62
4-Bromofluorobenzene
Concen: 48.657 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD076857.D
Acq: 28 Jul 2023 16:22

Tgt Ion: 95 Resp: 119765
Ion Ratio Lower Upper
95 100
174 89.0 0.0 201.8
176 84.0 0.0 191.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 117
Delta R.T. -0.000 min
Lab File: VD076857.D
Acq: 28 Jul 2023 16:22

Instrument :
MSVOA_D
ClientSampleId :
B-P16B

Tgt Ion:117 Resp: 320803
Ion Ratio Lower Upper
117 100
82 54.0 38.3 57.5
119 32.4 25.0 37.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.006 min
Lab File: VD076857.D
Acq: 28 Jul 2023 16:22

Tgt Ion:152 Resp: 140402
Ion Ratio Lower Upper
152 100
115 57.9 27.1 81.3
150 155.2 0.0 353.6

