


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD072919\
Data File : VD063189.D
Acq On : 29 Jul 2019 8:50
Operator : VA/AP
Sample : BFB
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
Title : SW846 8260
Last Update : Tue Jul 30 03:04:50 2019



AutoFind: Scans 1231, 1232, 1233; Background Corrected with Scan 1226

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	52418	PASS
75	95	30	60	47.2	109725	PASS
95	95	100	100	100.0	232554	PASS
96	95	5	9	5.7	13370	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	91.9	213674	PASS
175	174	5	9	6.8	14443	PASS
176	174	95	101	98.0	209365	PASS
177	176	5	9	5.6	11659	PASS