


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD072920\
 Data File : VD066081.D
 Acq On : 29 Jul 2020 09:16
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072120S.M
 Title : SW846 8260
 Last Update : Tue Jul 21 12:34:13 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.1	20496	PASS
75	95	30	60	49.7	59618	PASS
95	95	100	100	100.0	119885	PASS
96	95	5	9	7.2	8599	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	91.4	109562	PASS
175	174	5	9	7.7	8437	PASS
176	174	95	101	98.7	108149	PASS
177	176	5	9	6.6	7127	PASS