

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072921\
 Data File : VD069916.D
 Acq On : 29 Jul 2021 17:26
 Operator : VA/SY
 Sample : M3210-05RE
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-5(1.0-5.0)RE

Quant Time: Jul 30 00:57:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	23734	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	39096	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	40925	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	17579	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	23681	86.756	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 173.520%#	
35) Dibromofluoromethane	7.908	113	14042	56.975	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.940%	
50) Toluene-d8	10.332	98	49590	52.181	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.360%	
62) 4-Bromofluorobenzene	12.620	95	14697	46.274	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 92.540%	
Target Compounds						Qvalue
16) Acetone	4.156	43	4624	93.775	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072921\
Data File : VD069916.D
Acq On : 29 Jul 2021 17:26
Operator : VA/SY
Sample : M3210-05RE
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-5(1.0-5.0)RE

Quant Time: Jul 30 00:57:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

