

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072921\
 Data File : VD069919.D
 Acq On : 29 Jul 2021 18:52
 Operator : VA/SY
 Sample : M3216-11
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SPLP-LEACHATE-AMENDED-COM-

Quant Time: Jul 30 00:58:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	546984	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1070044	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1111346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	434966	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	412203	65.525	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.040%
35) Dibromofluoromethane	7.914	113	347996	51.589	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.180%
50) Toluene-d8	10.332	98	1429321	54.952	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.900%
62) 4-Bromofluorobenzene	12.626	95	526859	60.609	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	121.220%
Target Compounds						
						Qvalue
16) Acetone	4.156	43	87747	77.214	ug/l	99
17) Carbon Disulfide	4.409	76	655801	29.307	ug/l	99
25) 2-Butanone	7.214	43	15167	9.842	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072921\
Data File : VD069919.D
Acq On : 29 Jul 2021 18:52
Operator : VA/SY
Sample : M3216-11
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SPLP-LEACHATE-AMENDED-COM-

Quant Time: Jul 30 00:58:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

