

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072921\
 Data File : VD069920.D
 Acq On : 29 Jul 2021 19:21
 Operator : VA/SY
 Sample : M3230-02
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ETGI-357

Quant Time: Jul 30 00:58:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	533690	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1041700	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1079771	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	455010	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	376885	61.403	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	122.800%	
35) Dibromofluoromethane	7.914	113	355319	54.108	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	108.220%	
50) Toluene-d8	10.338	98	1357342	53.604	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	107.200%	
62) 4-Bromofluorobenzene	12.626	95	476459	56.302	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	112.600%	
Target Compounds						Qvalue
20) Methylene Chloride	4.962	84	169914	16.986	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072921\
Data File : VD069920.D
Acq On : 29 Jul 2021 19:21
Operator : VA/SY
Sample : M3230-02
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ETGI-357

Quant Time: Jul 30 00:58:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

