

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD073019\
 Data File : VD063251.D
 Acq On : 30 Jul 2019 19:34
 Operator : VA/AP
 Sample : K4049-07RE
 Misc : 4.50µl/5ml/MSVOA D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-7RE

Quant Time: Jul 31 06:39:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:04:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	859163	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	1109307	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	789985	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	278645	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	356812	37.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	74.86%	
35) Dibromofluoromethane	6.91	113	363672	38.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	76.80%	
50) Toluene-d8	10.40	98	678915	32.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	64.06%	
62) 4-Bromofluorobenzene	13.55	95	218907	22.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	45.32%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD073019\
Data File : VD063251.D
Acq On : 30 Jul 2019 19:34
Operator : VA/AP
Sample : K4049-07RE
Misc : 4.50µl/5ml/MSVOA D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-7RE

Quant Time: Jul 31 06:39:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:04:50 2019
Response via : Initial Calibration

