


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD073020\
Data File : VD066100.D
Acq On : 30 Jul 2020 09:31
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072120S.M
Title : SW846 8260
Last Update : Tue Jul 21 12:34:13 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	19869	PASS
75	95	30	60	50.7	54453	PASS
95	95	100	100	100.0	107370	PASS
96	95	5	9	6.5	6980	PASS
173	174	0.00	2	0.5	458	PASS
174	95	50	100	93.4	100234	PASS
175	174	5	9	8.3	8356	PASS
176	174	95	101	95.8	96072	PASS
177	176	5	9	7.0	6737	PASS