

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073021\
 Data File : VD069936.D
 Acq On : 30 Jul 2021 15:00
 Operator : VA/SY
 Sample : M3238-03
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 221

Quant Time: Jul 31 00:38:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	12910	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	21370	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	19919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	6336	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	11893	80.100	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 160.200%	
35) Dibromofluoromethane	7.914	113	7079	52.547	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.100%	
50) Toluene-d8	10.338	98	25094	48.308	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 96.620%	
62) 4-Bromofluorobenzene	12.620	95	6904	39.769	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 79.540%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073021\
Data File : VD069936.D
Acq On : 30 Jul 2021 15:00
Operator : VA/SY
Sample : M3238-03
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
221

Quant Time: Jul 31 00:38:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

