

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073123\
 Data File : VD076881.D
 Acq On : 31 Jul 2023 18:21
 Operator : KP/SY
 Sample : 03842-03
 Misc : 5.01G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 350

Quant Time: Aug 01 02:13:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	165918	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	323711	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	307766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	134719	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	85387	51.668	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.340%
35) Dibromofluoromethane	7.804	113	105387	51.807	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.620%
50) Toluene-d8	10.269	98	403461	55.442	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	110.880%
62) 4-Bromofluorobenzene	12.575	95	111939	47.549	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073123\
Data File : VD076881.D
Acq On : 31 Jul 2023 18:21
Operator : KP/SY
Sample : 03842-03
Misc : 5.01G/5.0ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
350

Quant Time: Aug 01 02:13:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

