

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073123\
 Data File : VD076885.D
 Acq On : 31 Jul 2023 20:15
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 01 02:14:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.391	0.379	3.1	77	0.00
3 P	Chloromethane	0.674	0.682	-1.2	79	0.00
4 C	Vinyl Chloride	0.899	0.916	-1.9#	80	0.00
5 T	Bromomethane	0.740	0.663	10.4	74	0.00
6 T	Chloroethane	0.679	0.709	-4.4	82	0.00
7 T	Trichlorofluoromethane	0.800	0.821	-2.6	84	0.00
8 T	Diethyl Ether	0.275	0.274	0.4	72	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.526	0.565	-7.4	86	0.00
10 T	Methyl Iodide	0.593	0.589	0.7	73	0.00
11 T	Tert butyl alcohol	0.037	0.040	-8.1	85	0.02
12 CM	1,1-Dichloroethene	0.494	0.481	2.6#	77	0.00
13 T	Acrolein	0.025	0.024	4.0	83	0.00
14 T	Allyl chloride	0.649	0.651	-0.3	76	-0.01
15 T	Acrylonitrile	0.135	0.145	-7.4	81	0.00
16 T	Acetone	0.119	0.092	22.7	59	0.00
17 T	Carbon Disulfide	1.154	1.045	9.4	69	0.00
18 T	Methyl Acetate	0.447	0.501	-12.1	85	0.00
19 T	Methyl tert-butyl Ether	1.325	1.381	-4.2	78	0.00
20 T	Methylene Chloride	0.984	0.921	6.4	99	0.00
21 T	trans-1,2-Dichloroethene	0.574	0.585	-1.9	77	0.00
22 T	Diisopropyl ether	1.537	1.698	-10.5	83	0.00
23 T	Vinyl Acetate	0.760	0.809	-6.4	79	0.00
24 P	1,1-Dichloroethane	1.036	1.114	-7.5	82	0.00
25 T	2-Butanone	0.159	0.168	-5.7	78	0.00
26 T	2,2-Dichloropropane	0.893	0.862	3.5	76	0.00
27 T	cis-1,2-Dichloroethene	0.708	0.746	-5.4	81	0.00
28 T	Bromochloromethane	0.259	0.249	3.9	80	0.00
29 T	Tetrahydrofuran	0.094	0.104	-10.6	85	0.01
30 C	Chloroform	1.135	1.268	-11.7#	89	0.00
31 T	Cyclohexane	0.733	0.676	7.8	79	0.00
32 T	1,1,1-Trichloroethane	0.916	1.009	-10.2	85	0.00
33 S	1,2-Dichloroethane-d4	0.498	0.531	-6.6	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.314	0.352	-12.1	93	0.00
36 T	1,1-Dichloropropene	0.411	0.443	-7.8	83	0.00
37 T	Ethyl Acetate	0.184	0.224	-21.7	86	0.00
38 T	Carbon Tetrachloride	0.393	0.439	-11.7	86	0.00
39 T	Methylcyclohexane	0.474	0.473	0.2	76	0.00
40 TM	Benzene	1.295	1.390	-7.3	82	0.00
41 T	Methacrylonitrile	0.112	0.105	6.3	72	0.00
42 TM	1,2-Dichloroethane	0.337	0.386	-14.5	85	0.00
43 T	Isopropyl Acetate	0.358	0.399	-11.5	82	0.00
44 TM	Trichloroethene	0.353	0.366	-3.7	78	0.00
45 C	1,2-Dichloropropane	0.334	0.371	-11.1#	79	0.00
46 T	Dibromomethane	0.196	0.231	-17.9	84	0.00
47 T	Bromodichloromethane	0.491	0.557	-13.4	82	0.00
48 T	Methyl methacrylate	0.163	0.186	-14.1	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073123\
 Data File : VD076885.D
 Acq On : 31 Jul 2023 20:15
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 01 02:14:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	83	0.00
50 S	Toluene-d8	1.124	1.136	-1.1	83	0.00
51 T	4-Methyl-2-Pentanone	0.191	0.223	-16.8	83	0.00
52 CM	Toluene	0.851	0.903	-6.1#	78	0.00
53 T	t-1,3-Dichloropropene	0.456	0.491	-7.7	77	0.00
54 T	cis-1,3-Dichloropropene	0.537	0.574	-6.9	77	0.00
55 T	1,1,2-Trichloroethane	0.286	0.321	-12.2	81	0.00
56 T	Ethyl methacrylate	0.330	0.368	-11.5	77	0.00
57 T	1,3-Dichloropropane	0.465	0.516	-11.0	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.128	0.120	6.3	71	0.00
59 T	2-Hexanone	0.136	0.151	-11.0	78	0.00
60 T	Dibromochloromethane	0.332	0.388	-16.9	84	0.00
61 T	1,2-Dibromoethane	0.265	0.299	-12.8	80	0.00
62 S	4-Bromofluorobenzene	0.364	0.393	-8.0	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	72	0.00
64 T	Tetrachloroethene	0.285	0.306	-7.4	79	0.00
65 PM	Chlorobenzene	1.006	1.098	-9.1	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.418	-11.8	81	0.00
67 C	Ethyl Benzene	1.766	1.913	-8.3#	78	0.00
68 T	m/p-Xylenes	0.678	0.737	-8.7	78	0.00
69 T	o-Xylene	0.656	0.706	-7.6	77	0.00
70 T	Styrene	1.148	1.275	-11.1	79	0.00
71 P	Bromoform	0.205	0.243	-18.5	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	75	0.00
73 T	Isopropylbenzene	3.628	3.864	-6.5	80	0.00
74 T	N-amyl acetate	0.764	0.848	-11.0	83	0.00
75 P	1,1,2,2-Tetrachloroethane	0.779	0.906	-16.3	86	0.00
76 T	1,2,3-Trichloropropane	0.475	0.548	-15.4	84	0.00
77 T	Bromobenzene	0.837	0.925	-10.5	81	0.00
78 T	n-propylbenzene	4.398	4.806	-9.3	82	0.00
79 T	2-Chlorotoluene	2.411	2.590	-7.4	80	0.00
80 T	1,3,5-Trimethylbenzene	2.933	3.192	-8.8	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.223	0.235	-5.4	75	0.00
82 T	4-Chlorotoluene	2.519	2.717	-7.9	79	0.00
83 T	tert-Butylbenzene	2.583	2.802	-8.5	80	0.00
84 T	1,2,4-Trimethylbenzene	2.997	3.163	-5.5	77	0.00
85 T	sec-Butylbenzene	3.946	4.343	-10.1	82	0.00
86 T	p-Isopropyltoluene	3.259	3.561	-9.3	81	0.00
87 T	1,3-Dichlorobenzene	1.722	1.885	-9.5	81	0.00
88 T	1,4-Dichlorobenzene	1.687	1.779	-5.5	79	0.00
89 T	n-Butylbenzene	3.189	3.438	-7.8	81	0.00
90 T	Hexachloroethane	0.602	0.717	-19.1	87	0.00
91 T	1,2-Dichlorobenzene	1.518	1.678	-10.5	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.111	0.117	-5.4	82	0.00
93 T	1,2,4-Trichlorobenzene	0.927	0.988	-6.6	80	0.00
94 T	Hexachlorobutadiene	0.446	0.465	-4.3	79	0.00
95 T	Naphthalene	2.006	2.193	-9.3	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073123\
Data File : VD076885.D
Acq On : 31 Jul 2023 20:15
Operator : KP/SY
Sample : VSTDCCC050
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Aug 01 02:14:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.831	0.894	-7.6	81	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6