

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073123\
 Data File : VD076880.D
 Acq On : 31 Jul 2023 17:53
 Operator : KP/SY
 Sample : 03842-01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 342

Quant Time: Aug 01 02:13:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	166106	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	325079	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	310225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	138990	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	86094	52.037	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.080%
35) Dibromofluoromethane	7.805	113	107029	52.393	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.780%
50) Toluene-d8	10.269	98	403103	55.160	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	110.320%
62) 4-Bromofluorobenzene	12.575	95	117362	49.643	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.280%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073123\
Data File : VD076880.D
Acq On : 31 Jul 2023 17:53
Operator : KP/SY
Sample : 03842-01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
342


Quant Time: Aug 01 02:13:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD076880.D
Acq: 31 Jul 2023 17:53

Instrument :
MSVOA_D
ClientSampleId :
342







Tgt Ion:168 Resp: 166106
Ion Ratio Lower Upper
168 100
99 55.6 36.7 55.1#



#33
1,2-Dichloroethane-d4
Concen: 52.037 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. -0.000 min
Lab File: VD076880.D
Acq: 31 Jul 2023 17:53

Tgt Ion: 65 Resp: 86094
Ion Ratio Lower Upper
65 100
67 55.8 0.0 115.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1

Delta R.T. -0.000 min

Lab File: VD076880.D

Acq: 31 Jul 2023 17:53

Instrument :

MSVOA_D

ClientSampleId :

342





Tgt Ion:114 Resp: 325079

Ion Ratio Lower Upper

114 100

63 17.3 0.0 33.4

88 13.1 0.0 26.2



#35

Dibromofluoromethane

Concen: 52.393 ug/l

RT: 7.805 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VD076880.D

Acq: 31 Jul 2023 17:53




Tgt Ion:113 Resp: 107029


Ion Ratio Lower Upper

113 100

111 100.6 84.7 127.1

192 17.9 16.8 25.2









#50
Toluene-d8
Concen: 55.160 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD076880.D
Acq: 31 Jul 2023 17:53




Instrument :
MSVOA_D
ClientSampleId :
342


Tgt Ion: 98 Resp: 403103
Ion Ratio Lower Upper
98 100
100 65.3 50.3 75.5



#62
4-Bromofluorobenzene
Concen: 49.643 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD076880.D
Acq: 31 Jul 2023 17:53

Tgt Ion: 95 Resp: 117362
Ion Ratio Lower Upper
95 100
174 85.2 0.0 201.8
176 85.1 0.0 191.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD076880.D
Acq: 31 Jul 2023 17:53

Instrument :
MSVOA_D
ClientSampleId :
342

Tgt Ion:117 Resp: 310225
Ion Ratio Lower Upper
117 100
82 53.3 38.3 57.5
119 34.6 25.0 37.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.006 min
Lab File: VD076880.D
Acq: 31 Jul 2023 17:53

Tgt Ion:152 Resp: 138990
Ion Ratio Lower Upper
152 100
115 56.6 27.1 81.3
150 152.9 0.0 353.6

