

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073123\
 Data File : VD076883.D
 Acq On : 31 Jul 2023 19:19
 Operator : KP/SY
 Sample : 03840-02
 Misc : 7.36G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BUILDING-B-5-(5-10)

Quant Time: Aug 01 02:13:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	154775	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	312063	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	300974	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	128802	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	90516	58.715	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.440%
35) Dibromofluoromethane	7.805	113	104718	53.400	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.800%
50) Toluene-d8	10.269	98	381624	54.399	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.800%
62) 4-Bromofluorobenzene	12.575	95	112028	49.363	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.720%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD073123\
Data File : VD076883.D
Acq On : 31 Jul 2023 19:19
Operator : KP/SY
Sample : 03840-02
Misc : 7.36G/5.0ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BUILDING-B-5-(5-10)


Quant Time: Aug 01 02:13:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD076883.D
Acq: 31 Jul 2023 19:19

Instrument :
MSVOA_D
ClientSampleId :
BUILDING-B-5-(5-10)







Tgt Ion:168 Resp: 154775
Ion Ratio Lower Upper
168 100
99 56.2 36.7 55.1#



#33
1,2-Dichloroethane-d4
Concen: 58.715 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. -0.000 min
Lab File: VD076883.D
Acq: 31 Jul 2023 19:19

Tgt Ion: 65 Resp: 90516
Ion Ratio Lower Upper
65 100
67 54.6 0.0 115.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD076883.D

Acq: 31 Jul 2023 19:19

Instrument :

MSVOA_D

ClientSampleId :

BUILDING-B-5-(5-10)





Tgt Ion:114 Resp: 312063

Ion Ratio Lower Upper

114 100

63 16.0 0.0 33.4

88 12.9 0.0 26.2



#35

Dibromofluoromethane

Concen: 53.400 ug/l

RT: 7.805 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VD076883.D

Acq: 31 Jul 2023 19:19




Tgt Ion:113 Resp: 104718


Ion Ratio Lower Upper

113 100

111 99.9 84.7 127.1

192 17.0 16.8 25.2









#50
Toluene-d8
Concen: 54.399 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD076883.D
Acq: 31 Jul 2023 19:19




Instrument :
MSVOA_D
ClientSampleId :
BUILDING-B-5-(5-10)


Tgt Ion: 98 Resp: 381624
Ion Ratio Lower Upper
98 100
100 64.8 50.3 75.5



#62
4-Bromofluorobenzene
Concen: 49.363 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD076883.D
Acq: 31 Jul 2023 19:19

Tgt Ion: 95 Resp: 112028
Ion Ratio Lower Upper
95 100
174 87.1 0.0 201.8
176 83.1 0.0 191.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD076883.D
Acq: 31 Jul 2023 19:19

Instrument :
MSVOA_D
ClientSampleId :
BUILDING-B-5-(5-10)

Tgt Ion:117 Resp: 300974
Ion Ratio Lower Upper
117 100
82 53.9 38.3 57.5
119 32.0 25.0 37.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.006 min
Lab File: VD076883.D
Acq: 31 Jul 2023 19:19

Tgt Ion:152 Resp: 128802
Ion Ratio Lower Upper
152 100
115 55.7 27.1 81.3
150 157.8 0.0 353.6

