

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD080119\
 Data File : VD063286.D
 Acq On : 1 Aug 2019 10:52
 Operator : VA/AP
 Sample : VD0801SBL01
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0801SBL01

Quant Time: Aug 02 07:33:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:04:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	1107103	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.20	114	1460430	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	1229448	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	639578	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	576879	46.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.92%	
35) Dibromofluoromethane	6.91	113	619073	49.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.30%	
50) Toluene-d8	10.40	98	1347728	48.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.60%	
62) 4-Bromofluorobenzene	13.55	95	566736	44.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.14%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD080119\
Data File : VD063286.D
Acq On : 1 Aug 2019 10:52
Operator : VA/AP
Sample : VD0801SBL01
Misc : 5.00g/5ml/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0801SBL01

Quant Time: Aug 02 07:33:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:04:50 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.03 min Scan# 569

Delta R.T. -0.00 min

Lab File: VD063286.D

Acq: 1 Aug 2019 10:52

Instrument :

MSVOA_D

ClientSampleId :


VD0801SBL01

Tot Ion:168 Resp: 1107103

Ion Ratio Lower Upper




168 100

99 58.0 43.7 65.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.95 (98.65 to 99.65): VDC



#33

1,2-Dichloroethane-d4

Concen: 46.960 ug/l

RT: 7.44 min Scan# 610

Delta R.T. -0.00 min

Lab File: VD063286.D


Acq: 1 Aug 2019 10:52

Tgt Ion: 65 Resp: 576879

Ion Ratio Lower Upper



65 100


67 43.4 0.0 85.8



Abundance Ion 64.95 (64.65 to 65.65): VDC

Ion 66.90 (66.60 to 67.60): VDC









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.20 min Scan# 688
 Delta R.T. -0.00 min
 Lab File: VD063286.D
 Acq: 1 Aug 2019 10:52

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0801SBL01

Tot Ion:114 Resp: 1460430
 Ion Ratio Lower Upper
 114 100
 63 19.0 0.0 35.2
 88 18.7 0.0 34.6




Abundance Ion 114.00 (113.70 to 114.70): V
 Ion 62.95 (62.65 to 63.65): V
 Ion 87.90 (87.60 to 88.60): V






#35
 Dibromofluoromethane
 Concen: 49.649 ug/l
 RT: 6.91 min Scan# 557
 Delta R.T. 0.01 min
 Lab File: VD063286.D
 Acq: 1 Aug 2019 10:52

Tgt Ion:113 Resp: 619073
 Ion Ratio Lower Upper
 113 100
 111 102.3 83.2 124.8
 192 16.3 11.8 17.8



Abundance Ion 112.90 (112.60 to 113.60): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 191.80 (191.50 to 192.50): V





#50

Toluene-d8

Concen: 48.296 ug/l

RT: 10.40 min Scan# 911

Delta R.T. -0.00 min

Lab File: VD063286.D

Acq: 1 Aug 2019 10:52

Instrument :

MSVOA_D

ClientSampleId :


VD0801SBL01

Tot Ion: 98 Resp: 1347728

Ion Ratio Lower Upper




98 100

100 70.2 53.2 79.8



Abundance Ion 98.05 (97.75 to 98.75): VDC

Ion 100.05 (99.75 to 100.75): VDC



#62

4-Bromofluorobenzene

Concen: 44.567 ug/l

RT: 13.55 min Scan# 1231

Delta R.T. -0.00 min

Lab File: VD063286.D

Acq: 1 Aug 2019 10:52


Tgt Ion: 95 Resp: 566736

Ion Ratio Lower Upper

95 100

174 92.4 0.0 205.4



176 88.3 0.0 211.8




Abundance Ion 95.00 (94.70 to 95.70): VDC

Ion 173.90 (173.60 to 174.60): VDC

Ion 175.90 (175.60 to 176.60): VDC









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 12.36 min Scan# 1110
Delta R.T. -0.00 min
Lab File: VD063286.D
Acq: 1 Aug 2019 10:52

Instrument :
MSVOA_D
ClientSampleId :
VD0801SBL01

Tot Ion:117 Resp: 1229448
Ion Ratio Lower Upper
117 100
82 50.4 40.1 60.1
119 33.7 26.5 39.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 14.51 min Scan# 1329
Delta R.T. -0.00 min
Lab File: VD063286.D
Acq: 1 Aug 2019 10:52

Tgt Ion:152 Resp: 639578
Ion Ratio Lower Upper
152 100
115 55.7 27.5 82.4
150 155.6 0.0 320.6

