

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD080219\
 Data File : VD063322.D
 Acq On : 2 Aug 2019 17:44
 Operator : VA/AP
 Sample : K4130-03
 Misc : 6.15a/5ml/MSVOA D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 1S-F2-(0-1)

Quant Time: Aug 04 04:28:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:04:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	1178469	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.20	114	1585669	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	1275163	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	592118	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.43	65	510457	39.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.08%	
35) Dibromofluoromethane	6.91	113	616615	45.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.10%	
50) Toluene-d8	10.40	98	1368820	45.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.36%	
62) 4-Bromofluorobenzene	13.55	95	519450	37.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.24%	
Target Compounds						
20) Methylene Chloride	3.79	84	29939	2.907	ug/l #	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD080219\
Data File : VD063322.D
Acq On : 2 Aug 2019 17:44
Operator : VA/AP
Sample : K4130-03
Misc : 6.15g/5ml/MSVOA D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
1S-F2-(0-1)

Quant Time: Aug 04 04:28:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:04:50 2019
Response via : Initial Calibration

