

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD080219\
 Data File : VD063323.D
 Acq On : 2 Aug 2019 18:12
 Operator : VA/AP
 Sample : K4130-04
 Misc : 4.12g/5ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 1S-F3-(0-1)

Quant Time: Aug 04 04:29:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:04:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	1097983	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	1456957	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	1120173	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	589737	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	487845	40.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	80.08%	
35) Dibromofluoromethane	6.91	113	550804	44.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.56%	
50) Toluene-d8	10.39	98	1187471	42.65	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	85.30%	
62) 4-Bromofluorobenzene	13.54	95	481861	37.98	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	75.96%	
Target Compounds						
20) Methylene Chloride	3.80	84	40460	4.217	ug/l	89
52) Toluene	10.49	92	32958	1.778	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD080219\
Data File : VD063323.D
Acq On : 2 Aug 2019 18:12
Operator : VA/AP
Sample : K4130-04
Misc : 4.12g/5ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
1S-F3-(0-1)

Quant Time: Aug 04 04:29:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:04:50 2019
Response via : Initial Calibration

