

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD080219\
 Data File : VD063342.D
 Acq On : 3 Aug 2019 4:38
 Operator : VA/AP
 Sample : K4129-02
 Misc : 5.34g/5ml/MSVOA D/SOIL
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 1S-H-(30)

Quant Time: Aug 04 05:05:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:04:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	687851	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.22	114	926751	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	822993	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	378826	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	427929	56.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.14%	
35) Dibromofluoromethane	6.92	113	432010	54.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.20%	
50) Toluene-d8	10.40	98	846173	47.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.56%	
62) 4-Bromofluorobenzene	13.55	95	344409	42.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.36%	
Target Compounds						
16) Acetone	3.22	43	29268	19.681	ug/l	99
20) Methylene Chloride	3.81	84	22676	3.772	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD080219\
Data File : VD063342.D
Acq On : 3 Aug 2019 4:38
Operator : VA/AP
Sample : K4129-02
Misc : 5.34g/5ml/MSVOA D/SOIL
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
1S-H-(30)

Quant Time: Aug 04 05:05:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:04:50 2019
Response via : Initial Calibration

