

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080221\
 Data File : VD069950.D
 Acq On : 02 Aug 2021 12:19
 Operator : VA/SY
 Sample : M3253-03
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-2-COMP

Quant Time: Aug 03 02:55:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	9392	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	16554	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	14136	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	5574	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	7889	73.035	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	146.080%
35) Dibromofluoromethane	7.914	113	5363	51.391	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.780%
50) Toluene-d8	10.338	98	18855	46.857	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.720%
62) 4-Bromofluorobenzene	12.626	95	5359	39.850	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	79.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080221\
Data File : VD069950.D
Acq On : 02 Aug 2021 12:19
Operator : VA/SY
Sample : M3253-03
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-2-COMP

Quant Time: Aug 03 02:55:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

