

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080221\
 Data File : VD069943.D
 Acq On : 02 Aug 2021 09:00
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 03 02:52:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.526	0.470	10.6	94	0.00
3 P	Chloromethane	0.763	0.660	13.5	92	0.00
4 C	Vinyl Chloride	0.803	0.740	7.8#	94	0.00
5 T	Bromomethane	0.526	0.500	4.9	103	0.00
6 T	Chloroethane	0.520	0.493	5.2	96	0.00
7 T	Trichlorofluoromethane	0.991	0.969	2.2	100	0.00
8 T	Diethyl Ether	0.279	0.257	7.9	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.630	0.625	0.8	103	0.01
10 T	Methyl Iodide	0.550	0.601	-9.3	95	0.00
11 T	Tert butyl alcohol	0.030	0.025	16.7	89	-0.01
12 CM	1,1-Dichloroethene	0.572	0.557	2.6#	95	0.00
13 T	Acrolein	0.042	0.036	14.3	89	0.00
14 T	Allyl chloride	0.839	0.838	0.1	95	0.00
15 T	Acrylonitrile	0.131	0.126	3.8	96	0.00
16 T	Acetone	0.104	0.125	-20.2	113	0.00
17 T	Carbon Disulfide	2.045	1.923	6.0	94	0.00
18 T	Methyl Acetate	0.301	0.273	9.3	98	0.00
19 T	Methyl tert-butyl Ether	1.185	1.176	0.8	95	0.00
20 T	Methylene Chloride	0.946	0.760	19.7	96	0.00
21 T	trans-1,2-Dichloroethene	0.674	0.675	-0.1	97	0.00
22 T	Diisopropyl ether	1.823	1.883	-3.3	96	0.00
23 T	Vinyl Acetate	0.873	0.922	-5.6	95	0.00
24 P	1,1-Dichloroethane	1.259	1.220	3.1	97	0.00
25 T	2-Butanone	0.141	0.147	-4.3	101	0.00
26 T	2,2-Dichloropropane	0.957	1.029	-7.5	106	0.00
27 T	cis-1,2-Dichloroethene	0.719	0.711	1.1	97	0.00
28 T	Bromochloromethane	0.478	0.454	5.0	92	0.00
29 T	Tetrahydrofuran	0.091	0.091	0.0	94	0.00
30 C	Chloroform	1.266	1.232	2.7#	98	0.00
31 T	Cyclohexane	1.124	1.118	0.5	100	0.00
32 T	1,1,1-Trichloroethane	1.048	1.066	-1.7	101	0.00
33 S	1,2-Dichloroethane-d4	0.575	0.537	6.6	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.315	0.330	-4.8	100	0.00
36 T	1,1-Dichloropropene	0.540	0.570	-5.6	98	0.00
37 T	Ethyl Acetate	0.196	0.199	-1.5	95	0.00
38 T	Carbon Tetrachloride	0.493	0.541	-9.7	103	0.00
39 T	Methylcyclohexane	0.618	0.688	-11.3	99	0.00
40 TM	Benzene	1.563	1.642	-5.1	98	0.00
41 T	Methacrylonitrile	0.103	0.117	-13.6	111	0.00
42 TM	1,2-Dichloroethane	0.413	0.414	-0.2	95	0.00
43 T	Isopropyl Acetate	0.352	0.358	-1.7	96	0.00
44 TM	Trichloroethene	0.385	0.387	-0.5	96	0.00
45 C	1,2-Dichloropropane	0.411	0.427	-3.9#	99	0.00
46 T	Dibromomethane	0.202	0.208	-3.0	99	0.00
47 T	Bromodichloromethane	0.524	0.537	-2.5	97	0.00
48 T	Methyl methacrylate	0.169	0.184	-8.9	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080221\
 Data File : VD069943.D
 Acq On : 02 Aug 2021 09:00
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 03 02:52:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	97	0.00
50 S	Toluene-d8	1.215	1.295	-6.6	101	0.00
51 T	4-Methyl-2-Pentanone	0.183	0.199	-8.7	98	0.00
52 CM	Toluene	0.948	1.031	-8.8#	99	0.00
53 T	t-1,3-Dichloropropene	0.440	0.480	-9.1	104	0.00
54 T	cis-1,3-Dichloropropene	0.533	0.578	-8.4	104	0.00
55 T	1,1,2-Trichloroethane	0.285	0.292	-2.5	99	0.00
56 T	Ethyl methacrylate	0.312	0.341	-9.3	98	0.00
57 T	1,3-Dichloropropane	0.482	0.509	-5.6	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.134	0.153	-14.2	96	0.00
59 T	2-Hexanone	0.121	0.144	-19.0	105	0.00
60 T	Dibromochloromethane	0.323	0.340	-5.3	101	0.00
61 T	1,2-Dibromoethane	0.257	0.263	-2.3	97	0.00
62 S	4-Bromofluorobenzene	0.406	0.436	-7.4	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.319	0.330	-3.4	97	0.00
65 PM	Chlorobenzene	1.036	1.088	-5.0	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.384	-2.9	97	0.00
67 C	Ethyl Benzene	1.824	2.011	-10.3#	99	0.00
68 T	m/p-Xylenes	0.707	0.796	-12.6	100	0.00
69 T	o-Xylene	0.643	0.702	-9.2	98	0.00
70 T	Styrene	1.114	1.268	-13.8	100	0.00
71 P	Bromoform	0.181	0.185	-2.2	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.593	4.110	-14.4	101	0.00
74 T	N-amyl acetate	0.739	0.793	-7.3	98	0.00
75 P	1,1,2,2-Tetrachloroethane	0.717	0.738	-2.9	100	0.00
76 T	1,2,3-Trichloropropane	0.531	0.535	-0.8	105	0.00
77 T	Bromobenzene	0.809	0.839	-3.7	97	0.00
78 T	n-propylbenzene	4.614	5.361	-16.2	102	0.00
79 T	2-Chlorotoluene	2.680	2.914	-8.7	99	0.00
80 T	1,3,5-Trimethylbenzene	3.073	3.512	-14.3	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.175	0.211	-20.6	109	0.00
82 T	4-Chlorotoluene	2.802	3.099	-10.6	100	0.00
83 T	tert-Butylbenzene	2.499	2.928	-17.2	104	0.00
84 T	1,2,4-Trimethylbenzene	3.032	3.456	-14.0	101	0.00
85 T	sec-Butylbenzene	3.950	4.586	-16.1	103	0.00
86 T	p-Isopropyltoluene	3.209	3.746	-16.7	102	0.00
87 T	1,3-Dichlorobenzene	1.699	1.805	-6.2	101	0.00
88 T	1,4-Dichlorobenzene	1.693	1.751	-3.4	100	0.00
89 T	n-Butylbenzene	3.167	3.692	-16.6	103	0.00
90 T	Hexachloroethane	0.680	0.736	-8.2	104	0.00
91 T	1,2-Dichlorobenzene	1.466	1.524	-4.0	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.103	0.100	2.9	92	0.00
93 T	1,2,4-Trichlorobenzene	0.791	0.849	-7.3	99	0.00
94 T	Hexachlorobutadiene	0.484	0.530	-9.5	102	0.00
95 T	Naphthalene	1.412	1.514	-7.2	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080221\
Data File : VD069943.D
Acq On : 02 Aug 2021 09:00
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Aug 03 02:52:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.684	0.732	-7.0	98	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6