

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080222\
 Data File : VD073930.D
 Acq On : 02 Aug 2022 16:57
 Operator : VA/SY
 Sample : N3982-01
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-3-072922

Quant Time: Aug 03 04:13:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 02 05:13:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	102812	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	228159	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	232796	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	109491	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	71952	88.626	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	177.260%#
35) Dibromofluoromethane	7.909	113	83038	58.577	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	117.160%
50) Toluene-d8	10.326	98	294285	61.629	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	123.260%
62) 4-Bromofluorobenzene	12.620	95	102609	58.433	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080222\
Data File : VD073930.D
Acq On : 02 Aug 2022 16:57
Operator : VA/SY
Sample : N3982-01
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-3-072922

Quant Time: Aug 03 04:13:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 02 05:13:57 2022
Response via : Initial Calibration

