

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080223\
 Data File : VD076904.D
 Acq On : 02 Aug 2023 18:06
 Operator : KP/SY
 Sample : 03886-01
 Misc : 5.01G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 500-1A-1B

Quant Time: Aug 03 01:25:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	155158	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	307084	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	294287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	126305	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	82993	53.703	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.400%
35) Dibromofluoromethane	7.804	113	99681	51.655	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.320%
50) Toluene-d8	10.269	98	385615	55.859	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	111.720%
62) 4-Bromofluorobenzene	12.569	95	109317	48.950	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080223\
Data File : VD076904.D
Acq On : 02 Aug 2023 18:06
Operator : KP/SY
Sample : 03886-01
Misc : 5.01G/5.0ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
500-1A-1B

Quant Time: Aug 03 01:25:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

