

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080223\
 Data File : VD076905.D
 Acq On : 02 Aug 2023 18:34
 Operator : KP/SY
 Sample : 03864-01
 Misc : 5.02G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-1

Quant Time: Aug 03 01:25:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	91250	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	184393	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	173598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	57703	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	62971	69.284	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	138.560%
35) Dibromofluoromethane	7.804	113	66846	57.689	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.380%
50) Toluene-d8	10.269	98	220909	53.292	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.580%
62) 4-Bromofluorobenzene	12.575	95	60101	44.818	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080223\
Data File : VD076905.D
Acq On : 02 Aug 2023 18:34
Operator : KP/SY
Sample : 03864-01
Misc : 5.02G/5.0ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-1

Quant Time: Aug 03 01:25:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

