

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD080319\
 Data File : VD063354.D
 Acq On : 3 Aug 2019 11:50
 Operator : VA/AP
 Sample : VD0803SBL01
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0803SBL01

Quant Time: Aug 04 06:16:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:04:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	1100107	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	1522548	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	1294578	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	621516	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	528622	43.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.62%	
35) Dibromofluoromethane	6.92	113	623435	47.96	ug/l	0.02
Spiked Amount	50.000		Recovery	=	95.92%	
50) Toluene-d8	10.41	98	1355782	46.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.20%	
62) 4-Bromofluorobenzene	13.56	95	591679	44.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.26%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD080319\
Data File : VD063354.D
Acq On : 3 Aug 2019 11:50
Operator : VA/AP
Sample : VD0803SBL01
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0803SBL01

Quant Time: Aug 04 06:16:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:04:50 2019
Response via : Initial Calibration

