

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD080319\
 Data File : VD063369.D
 Acq On : 3 Aug 2019 18:44
 Operator : VA/AP
 Sample : K4135-07
 Misc : 7.57g/5ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 1S-A1-(30)

Quant Time: Aug 04 06:47:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:04:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	856653	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	8.22	114	1113657	50.00	ug/l	0.01
63) Chlorobenzene-d5	12.37	117	870484	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	14.52	152	362036	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	474401	49.91	ug/l	0.01
Spiked Amount 50.000			Recovery	=	99.82%	
35) Dibromofluoromethane	6.92	113	504499	53.06	ug/l	0.01
Spiked Amount 50.000			Recovery	=	106.12%	
50) Toluene-d8	10.41	98	1042986	49.01	ug/l	0.01
Spiked Amount 50.000			Recovery	=	98.02%	
62) 4-Bromofluorobenzene	13.55	95	377858	38.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.94%	
Target Compounds						
16) Acetone	3.22	43	39487	21.321	ug/l	98
20) Methylene Chloride	3.81	84	16574	2.214	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD080319\
Data File : VD063369.D
Acq On : 3 Aug 2019 18:44
Operator : VA/AP
Sample : K4135-07
Misc : 7.57g/5ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
1S-A1-(30)

Quant Time: Aug 04 06:47:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:04:50 2019
Response via : Initial Calibration

