

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080321\
 Data File : VD069972.D
 Acq On : 03 Aug 2021 18:05
 Operator : VA/SY
 Sample : M3268-16
 Misc : 8.41G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20210508-94

Quant Time: Aug 04 01:33:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	561565	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1131180	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	955492	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	302868	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	365091	56.529	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.060%
35) Dibromofluoromethane	7.914	113	368300	51.648	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.300%
50) Toluene-d8	10.338	98	1414193	51.432	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.860%
62) 4-Bromofluorobenzene	12.626	95	381144	41.476	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.960%
Target Compounds						Qvalue
16) Acetone	4.161	43	99587	85.358	ug/l	99
17) Carbon Disulfide	4.408	76	48463	2.110	ug/l	97
25) 2-Butanone	7.208	43	34339	21.703	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080321\
Data File : VD069972.D
Acq On : 03 Aug 2021 18:05
Operator : VA/SY
Sample : M3268-16
Misc : 8.41G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20210508-94

Quant Time: Aug 04 01:33:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

