

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080322\
 Data File : VD073943.D
 Acq On : 03 Aug 2022 16:31
 Operator : VA/SY
 Sample : N3996-01
 Misc : 5.97G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-CS

Quant Time: Aug 04 02:52:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 02 05:13:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	130342	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	223129	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	204733	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	87430	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	63177	59.819	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.640%
35) Dibromofluoromethane	7.908	113	68097	49.120	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.240%
50) Toluene-d8	10.332	98	195738	41.915	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	83.840%
62) 4-Bromofluorobenzene	12.620	95	69943	40.728	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080322\
Data File : VD073943.D
Acq On : 03 Aug 2022 16:31
Operator : VA/SY
Sample : N3996-01
Misc : 5.97G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-CS

Quant Time: Aug 04 02:52:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 02 05:13:57 2022
Response via : Initial Calibration

