

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080322\
 Data File : VD073946.D
 Acq On : 03 Aug 2022 18:01
 Operator : VA/SY
 Sample : N3947-02
 Misc : 5.07G/5.02ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NSB07-14-15

Quant Time: Aug 04 02:53:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 02 05:13:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	34754	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	66123	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	73478	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	34927	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	34404	127.468	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 254.940%#	
35) Dibromofluoromethane	7.903	113	27131	66.039	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 132.080%	
50) Toluene-d8	10.332	98	64014	46.257	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 92.520%	
62) 4-Bromofluorobenzene	12.620	95	28384	55.774	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 111.540%	
Target Compounds						
						Qvalue
16) Acetone	4.167	43	4008	63.467	ug/l	91
20) Methylene Chloride	4.950	84	15673	32.743	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080322\
Data File : VD073946.D
Acq On : 03 Aug 2022 18:01
Operator : VA/SY
Sample : N3947-02
Misc : 5.07G/5.02ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NSB07-14-15

Quant Time: Aug 04 02:53:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 02 05:13:57 2022
Response via : Initial Calibration

