



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080322\
 Data File : VD073937.D
 Acq On : 03 Aug 2022 09:41
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
 Title : SW846 8260
 Last Update : Tue Aug 02 05:13:57 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	9623	PASS
75	95	30	60	54.5	26760	PASS
95	95	100	100	100.0	49072	PASS
96	95	5	9	6.7	3306	PASS
173	174	0.00	2	1.4	544	PASS
174	95	50	100	82.1	40283	PASS
175	174	5	9	7.7	3117	PASS
176	174	95	101	95.4	38419	PASS
177	176	5	9	7.6	2931	PASS