

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080421\
 Data File : VD069984.D
 Acq On : 04 Aug 2021 14:18
 Operator : VA/SY
 Sample : M3268-09
 Misc : 9.78G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20210513-99

Quant Time: Aug 05 03:10:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	629320	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1234199	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1035425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	317657	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	368660	50.936	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 101.880%			
35) Dibromofluoromethane	7.914	113	390568	50.199	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 100.400%			
50) Toluene-d8	10.338	98	1544515	51.483	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 102.960%			
62) 4-Bromofluorobenzene	12.626	95	421795	42.069	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 84.140%			
Target Compounds						
16) Acetone	4.156	43	106568	81.507	ug/l	96
17) Carbon Disulfide	4.414	76	50553	1.964	ug/l	97
25) 2-Butanone	7.208	43	37802	21.320	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080421\
Data File : VD069984.D
Acq On : 04 Aug 2021 14:18
Operator : VA/SY
Sample : M3268-09
Misc : 9.78G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20210513-99

Quant Time: Aug 05 03:10:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

