

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080421\
 Data File : VD069987.D
 Acq On : 04 Aug 2021 15:44
 Operator : VA/SY
 Sample : M3279-13
 Misc : 7.80G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20210451-37

Quant Time: Aug 05 03:11:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	587632	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1153936	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	991623	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	319822	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	384495	56.892	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.780%
35) Dibromofluoromethane	7.914	113	381480	52.441	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.880%
50) Toluene-d8	10.338	98	1431430	51.032	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.060%
62) 4-Bromofluorobenzene	12.626	95	396153	42.260	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.520%
Target Compounds						
						Qvalue
16) Acetone	4.156	43	172813	141.550	ug/l	91
17) Carbon Disulfide	4.415	76	80070	3.331	ug/l #	94
25) 2-Butanone	7.214	43	48085	29.043	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080421\
Data File : VD069987.D
Acq On : 04 Aug 2021 15:44
Operator : VA/SY
Sample : M3279-13
Misc : 7.80G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20210451-37

Quant Time: Aug 05 03:11:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

