

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080421\
 Data File : VD069989.D
 Acq On : 04 Aug 2021 16:41
 Operator : VA/SY
 Sample : M3279-15
 Misc : 8.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20210491-77

Quant Time: Aug 05 03:12:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	556342	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1095434	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	919410	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	279132	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	351437	54.925	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery	= 109.860%		
35) Dibromofluoromethane	7.914	113	359746	52.095	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery	= 104.180%		
50) Toluene-d8	10.338	98	1362820	51.181	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery	= 102.360%		
62) 4-Bromofluorobenzene	12.626	95	366850	41.224	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery	= 82.440%		
Target Compounds						Qvalue
16) Acetone	4.162	43	94327	81.608	ug/l	91
17) Carbon Disulfide	4.409	76	41476	1.822	ug/l	100
25) 2-Butanone	7.214	43	32685	20.852	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080421\
Data File : VD069989.D
Acq On : 04 Aug 2021 16:41
Operator : VA/SY
Sample : M3279-15
Misc : 8.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20210491-77

Quant Time: Aug 05 03:12:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

