

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080421\
 Data File : VD069991.D
 Acq On : 04 Aug 2021 17:38
 Operator : VA/SY
 Sample : M3279-09
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20210589-AMENDED-COM-3

Quant Time: Aug 05 03:12:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	611893	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1206286	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	1082636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	435614	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	393893	55.972	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.940%
35) Dibromofluoromethane	7.914	113	367210	48.289	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.580%
50) Toluene-d8	10.337	98	1543604	52.643	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.280%
62) 4-Bromofluorobenzene	12.625	95	485970	49.591	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.180%
Target Compounds						
						Qvalue
16) Acetone	4.161	43	26311	20.697	ug/l	91
17) Carbon Disulfide	4.408	76	173147	6.917	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080421\
Data File : VD069991.D
Acq On : 04 Aug 2021 17:38
Operator : VA/SY
Sample : M3279-09
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20210589-AMENDED-COM-3

Quant Time: Aug 05 03:12:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

