

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080421\
 Data File : VD069977.D
 Acq On : 04 Aug 2021 10:55
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 05 03:08:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.526	0.439	16.5	84	0.00
3 P	Chloromethane	0.763	0.631	17.3	84	0.00
4 C	Vinyl Chloride	0.803	0.720	10.3#	87	0.00
5 T	Bromomethane	0.526	0.465	11.6	92	0.00
6 T	Chloroethane	0.520	0.468	10.0	87	0.00
7 T	Trichlorofluoromethane	0.991	0.907	8.5	90	0.00
8 T	Diethyl Ether	0.279	0.271	2.9	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.630	0.591	6.2	93	0.00
10 T	Methyl Iodide	0.550	0.575	-4.5	87	-0.01
11 T	Tert butyl alcohol	0.030	0.029	3.3	100	-0.01
12 CM	1,1-Dichloroethene	0.572	0.545	4.7#	89	0.00
13 T	Acrolein	0.042	0.045	-7.1	106	0.00
14 T	Allyl chloride	0.839	0.842	-0.4	92	0.00
15 T	Acrylonitrile	0.131	0.131	0.0	96	0.00
16 T	Acetone	0.104	0.121	-16.3	105	0.00
17 T	Carbon Disulfide	2.045	1.778	13.1	83	0.00
18 T	Methyl Acetate	0.301	0.274	9.0	95	0.00
19 T	Methyl tert-butyl Ether	1.185	1.220	-3.0	94	0.00
20 T	Methylene Chloride	0.946	0.795	16.0	96	0.00
21 T	trans-1,2-Dichloroethene	0.674	0.632	6.2	87	0.00
22 T	Diisopropyl ether	1.823	1.871	-2.6	91	0.00
23 T	Vinyl Acetate	0.873	0.960	-10.0	95	0.00
24 P	1,1-Dichloroethane	1.259	1.172	6.9	89	0.00
25 T	2-Butanone	0.141	0.150	-6.4	99	0.00
26 T	2,2-Dichloropropane	0.957	0.990	-3.4	98	0.00
27 T	cis-1,2-Dichloroethene	0.719	0.695	3.3	90	0.00
28 T	Bromochloromethane	0.478	0.521	-9.0	101	0.00
29 T	Tetrahydrofuran	0.091	0.098	-7.7	97	0.00
30 C	Chloroform	1.266	1.167	7.8#	89	0.00
31 T	Cyclohexane	1.124	1.072	4.6	92	0.00
32 T	1,1,1-Trichloroethane	1.048	0.925	11.7	84	0.00
33 S	1,2-Dichloroethane-d4	0.575	0.551	4.2	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.315	0.305	3.2	92	0.00
36 T	1,1-Dichloropropene	0.540	0.528	2.2	90	0.00
37 T	Ethyl Acetate	0.196	0.199	-1.5	95	0.00
38 T	Carbon Tetrachloride	0.493	0.490	0.6	93	0.00
39 T	Methylcyclohexane	0.618	0.622	-0.6	89	0.00
40 TM	Benzene	1.563	1.474	5.7	88	0.00
41 T	Methacrylonitrile	0.103	0.110	-6.8	103	0.00
42 TM	1,2-Dichloroethane	0.413	0.390	5.6	89	0.00
43 T	Isopropyl Acetate	0.352	0.363	-3.1	96	0.00
44 TM	Trichloroethene	0.385	0.338	12.2	83	0.00
45 C	1,2-Dichloropropane	0.411	0.395	3.9#	91	0.00
46 T	Dibromomethane	0.202	0.187	7.4	88	0.00
47 T	Bromodichloromethane	0.524	0.490	6.5	88	0.00
48 T	Methyl methacrylate	0.169	0.167	1.2	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080421\
 Data File : VD069977.D
 Acq On : 04 Aug 2021 10:55
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 05 03:08:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	94	0.00
50 S	Toluene-d8	1.215	1.288	-6.0	100	0.00
51 T	4-Methyl-2-Pentanone	0.183	0.200	-9.3	98	0.00
52 CM	Toluene	0.948	0.955	-0.7#	91	0.00
53 T	t-1,3-Dichloropropene	0.440	0.464	-5.5	100	0.00
54 T	cis-1,3-Dichloropropene	0.533	0.558	-4.7	100	0.00
55 T	1,1,2-Trichloroethane	0.285	0.271	4.9	91	0.00
56 T	Ethyl methacrylate	0.312	0.341	-9.3	98	0.00
57 T	1,3-Dichloropropane	0.482	0.484	-0.4	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.134	0.169	-26.1#	106	0.00
59 T	2-Hexanone	0.121	0.132	-9.1	95	0.00
60 T	Dibromochloromethane	0.323	0.314	2.8	92	0.00
61 T	1,2-Dibromoethane	0.257	0.255	0.8	94	0.00
62 S	4-Bromofluorobenzene	0.406	0.435	-7.1	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.319	0.306	4.1	87	0.00
65 PM	Chlorobenzene	1.036	1.030	0.6	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.366	1.9	90	0.00
67 C	Ethyl Benzene	1.824	1.915	-5.0#	91	0.00
68 T	m/p-Xylenes	0.707	0.743	-5.1	90	0.00
69 T	o-Xylene	0.643	0.676	-5.1	91	0.00
70 T	Styrene	1.114	1.206	-8.3	92	0.00
71 P	Bromoform	0.181	0.182	-0.6	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.593	3.792	-5.5	92	0.00
74 T	N-amyl acetate	0.739	0.788	-6.6	96	0.00
75 P	1,1,2,2-Tetrachloroethane	0.717	0.717	0.0	96	0.00
76 T	1,2,3-Trichloropropane	0.531	0.494	7.0	96	0.00
77 T	Bromobenzene	0.809	0.807	0.2	93	0.00
78 T	n-propylbenzene	4.614	4.843	-5.0	91	0.00
79 T	2-Chlorotoluene	2.680	2.723	-1.6	91	0.00
80 T	1,3,5-Trimethylbenzene	3.073	3.230	-5.1	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.175	0.206	-17.7	106	0.00
82 T	4-Chlorotoluene	2.802	2.836	-1.2	90	0.00
83 T	tert-Butylbenzene	2.499	2.637	-5.5	93	0.00
84 T	1,2,4-Trimethylbenzene	3.032	3.207	-5.8	93	0.00
85 T	sec-Butylbenzene	3.950	4.216	-6.7	93	0.00
86 T	p-Isopropyltoluene	3.209	3.401	-6.0	92	0.00
87 T	1,3-Dichlorobenzene	1.699	1.667	1.9	92	0.00
88 T	1,4-Dichlorobenzene	1.693	1.635	3.4	92	0.00
89 T	n-Butylbenzene	3.167	3.385	-6.9	94	0.00
90 T	Hexachloroethane	0.680	0.665	2.2	93	0.00
91 T	1,2-Dichlorobenzene	1.466	1.455	0.8	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.103	0.103	0.0	94	0.00
93 T	1,2,4-Trichlorobenzene	0.791	0.830	-4.9	96	0.00
94 T	Hexachlorobutadiene	0.484	0.485	-0.2	92	0.00
95 T	Naphthalene	1.412	1.544	-9.3	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080421\
Data File : VD069977.D
Acq On : 04 Aug 2021 10:55
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Aug 05 03:08:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.684	0.726	-6.1	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6