

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080422\
 Data File : VD073958.D
 Acq On : 04 Aug 2022 17:36
 Operator : VA/SY
 Sample : N3962-15
 Misc : 2.39G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RVE-200

Quant Time: Aug 05 07:54:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 02 05:13:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	120026	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	205829	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.631	117	171798	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	65024	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	60451	62.356	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.720%
35) Dibromofluoromethane	7.908	113	62654	48.992	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.980%
50) Toluene-d8	10.332	98	166405	38.629	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	77.260%
62) 4-Bromofluorobenzene	12.620	95	52767	33.309	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	66.620%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.955	84	24464	6.740	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080422\
Data File : VD073958.D
Acq On : 04 Aug 2022 17:36
Operator : VA/SY
Sample : N3962-15
Misc : 2.39G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RVE-200

Quant Time: Aug 05 07:54:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 02 05:13:57 2022
Response via : Initial Calibration

