

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080422\
 Data File : VD073959.D
 Acq On : 04 Aug 2022 18:05
 Operator : VA/SY
 Sample : N3962-19
 Misc : 3.71G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RVE-204

Quant Time: Aug 05 07:54:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 02 05:13:57 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	105142	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	186177	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.631	117	153613	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	54824	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	58566	69.502	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	139.000%	
35) Dibromofluoromethane	7.902	113	61937	53.544	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	107.080%	
50) Toluene-d8	10.326	98	152497	39.137	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	78.280%	
62) 4-Bromofluorobenzene	12.620	95	50229	35.054	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	70.100%	
Target Compounds						Qvalue
20) Methylene Chloride	4.961	84	25071	10.383	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080422\
Data File : VD073959.D
Acq On : 04 Aug 2022 18:05
Operator : VA/SY
Sample : N3962-19
Misc : 3.71G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RVE-204

Quant Time: Aug 05 07:54:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 02 05:13:57 2022
Response via : Initial Calibration

