

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080422\
 Data File : VD073961.D
 Acq On : 04 Aug 2022 19:01
 Operator : VA/SY
 Sample : N3773-18
 Misc : 6.45G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DL-3(14.5-15)

Quant Time: Aug 05 07:54:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 02 05:13:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	130164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	232671	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	212316	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	93269	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	60618	57.276	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.560%
35) Dibromofluoromethane	7.908	113	68133	47.131	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.260%
50) Toluene-d8	10.326	98	191415	39.309	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	78.620%
62) 4-Bromofluorobenzene	12.620	95	70442	39.337	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.680%
Target Compounds						
16) Acetone	4.162	43	4393	25.216	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080422\
Data File : VD073961.D
Acq On : 04 Aug 2022 19:01
Operator : VA/SY
Sample : N3773-18
Misc : 6.45G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DL-3(14.5-15)

Quant Time: Aug 05 07:54:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 02 05:13:57 2022
Response via : Initial Calibration

