



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080422\
Data File : VD073948.D
Acq On : 04 Aug 2022 08:56
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080122S.M
Title : SW846 8260
Last Update : Tue Aug 02 05:13:57 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	8942	PASS
75	95	30	60	54.7	26080	PASS
95	95	100	100	100.0	47707	PASS
96	95	5	9	6.5	3080	PASS
173	174	0.00	2	1.8	678	PASS
174	95	50	100	80.9	38613	PASS
175	174	5	9	8.5	3298	PASS
176	174	95	101	96.6	37312	PASS
177	176	5	9	6.6	2469	PASS