


Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD080618\
Data File : VD059684.D
Acq On : 6 Aug 2018 9:53
Operator : VA/AP
Sample : BFB
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080618S.M
Title : SW846 8260
Last Update : Wed Aug 08 02:03:19 2018



AutoFind: Scans 1188, 1189, 1190; Background Corrected with Scan 1184

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	26.0	18441	PASS
75	95	30	60	53.1	37666	PASS
95	95	100	100	100.0	70962	PASS
96	95	5	9	6.5	4595	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	82.2	58309	PASS
175	174	5	9	8.0	4636	PASS
176	174	95	101	97.6	56885	PASS
177	176	5	9	6.5	3725	PASS